

VC EMS TRAINING BULLETIN

Bulletin 12
Date: 09/20/11

IMPROVING QUALITY...CONTINUOUSLY



*Ventura County
Emergency Medical Services Agency*

**STORING EZ-IO POWER DRIVER IN THE VASCULAR ACCESS
PACK (YELLOW CASES) MAY CAUSE ACCIDENTAL
ACTIVATION OF DRIVER
RESULTING IN A DRAINED BATTERY***



RECOMMENDATION: Remove the trigger guard when using the Vascular Access Pack to prevent accident activation of the EZ-IO power driver.

**The compartment designated for driver storage is tight and has been known to cause activation of the driver from pressure on the trigger guard. Please remember to check driver function when checking your equipment to assure this has not occurred.*

VENTURA COUNTY EMS AGENCY

Phone: 805-981-5301

Fax: 805-981-5300

Web Page: www.vchca.org/ph/ems

